




R. M. Young Company
Barometric Pressure Sensor Calibration Record
Serial Number: BPA1729
Calibration Date: February 14, 2025



- LEGEND:
- 40°C
 - 25°C
 - 0°C
 - 25°C
 - 60°C

Sensor calibration checked from 500 to 1100 hPa in 100 hPa increments at 5 temperatures.
Pressure reference traceable to NIST.

Average Difference from Pressure Reference: -0.05 hPa
Standard Deviation: 0.02 hPa