



THE EPPLEY LABORATORY, INC.

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840

Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppley.com

Calibration Certificate

Instrument: Precision Spectral Pyranometer, Model PSP, Serial Number 30600F3

Procedure: This pyranometer was compared in Eppley's Integrating Hemisphere according to procedures described in *ISO 9847 Section 5.3.1* and Technical Procedure, TP01 of The Eppley Laboratory, Inc.'s Quality Assurance Manual on Calibrations.

Transfer Standard: Eppley Standard Precision Pyranometer, Model SPP, Serial Number 37501F3


Results:
Sensitivity: $S = 7.73 \mu\text{V} / \text{Wm}^{-2}$
Uncertainty: $U_{95} = \pm 0.91\%$ (95% confidence level, $k=2$)
Resistance: 730 Ω at 23°C

Date of Test: December 7, 2020

Traceability: This calibration is traceable to the World Radiation Reference (WRR) through comparisons with Eppley's AHF standard self-calibrating cavity pyrhemometers which participated in the Twelfth International Pyrhemometric Comparisons (IPC XII) at Davos, Switzerland in September-October 2015. Unless otherwise stated in the remarks section below or on the Sales Order, the results of this calibration are "AS FOUND / AS LEFT".

Due Date: Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy.

Customer: URI GSO / RMR Co
Narragansett, RI / Seattle, WA

In Charge of Test: 

Eppley SO: 65872

Date of Certificate: December 10, 2020

Remarks: