



THE EPPLEY LABORATORY, INC.

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840
Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

Calibration Certificate

Instrument: Eppley PSP Pyranometer #30600F3

Procedure: This pyranometer was compared in Eppley's Integrating Hemisphere at approximately 500-600 Wm^{-2} and a temperature of 23 ± 2 °C according to procedures described in *ISO 9847:1992 Section 5.3.1* and Technical Procedure, TP01 of The Eppley Laboratory, Inc.'s Quality Assurance Manual on Calibrations. The results relate only to the items tested, calibrated or sampled.

Calibration Results: **Sensitivity: $S = 7.76 \mu\text{V} / \text{Wm}^{-2}$**
Uncertainty: $U_{95} = \pm 1.05\%$ (95% confidence level, $k=2$)
Resistance: 731 Ω at 23°C


Date of Test: November 30, 2023

Transfer Standard: Eppley SPP Pyranometer #37501F3
which was calibrated against Eppley AHF Radiometer 14915 and Eppley 8-48 Pyranometer #29508 according to procedures described in ISO 9846:2023 between March 29 and May 1, 2023 with the Irradiance range of 459-978 Wm^{-2} , temperature range of 3-16 °C and a Solar Elevation range of 28-64°.

Traceability: This calibration is traceable to the World Radiation Reference (WRR) through comparisons with Eppley's AHF standard self-calibrating cavity pyrheliometers which participated in the Thirteenth International Pyrheliometric Comparisons (IPC XIII) at Davos, Switzerland in September-October 2021. Unless otherwise stated in the remarks section below or on the Sales Order, the results of this calibration are "AS FOUND / AS LEFT".

Due Date: Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy.

Customer: URI
Narragansett, RI

In Charge of Test: 

Eppley SO: 68138

Date of Certificate: December 4, 2023

Remarks: