




R. M. Young Company  
Barometric Pressure Sensor Calibration Record  
Serial Number: BPA1729  
Calibration Date: December 03, 2023



LEGEND:

- 40°C
- 25°C
- 0°C
- 25°C
- 60°C

Sensor calibration checked from 500 to 1100 hPa in 100 hPa increments at 5 temperatures.  
Pressure reference traceable to NIST.

Average Difference from Pressure Reference: -0.04 hPa  
Standard Deviation: 0.02 hPa