

Remote Measurements & Research Company
214 Euclid Av.
Seattle WA 98122
michael@rmrco.com



rad208cal_2010

rad201cal.2112

December 14, 2021

CALIBRATION RESULTS

Instrument	Radiometer Analog to Digital Interface (RAD)
Serial Number	201
RMR Job#	426
Customer	URI -Endeavor
Cal ID	2112
PSP	35149 (8.16)
PIR	35116 (3.80)
Lead Tech	J. Reynolds
Location	RMRCO Seattle
System recvd and inspected.	approx 12/2021
Shipped to customer.	approx 12/2021
Calibration Type	Electronic gain and ADC conversion.
After calibration	Returned to R/V Dyson R/V Endeavor

Same as RAD S/N 217:

System recvd and inspected	<10/21
PIR & PSP Cal at Eppley Labs	10/1-10/23/21
Electronic Cal	12/7/2021
Burn-In Test	12/8 - 12/10/2021
Est Ship Date	12/10/2021
Actual Ship Date	12/16/2021

DISCUSSION: Excellent calibration. The equipment has been handled well.

Program a1_radcal.m v3 Runtime:2021-12-14 (348) 08:37:26.0

Calibration date: 211212

SN: 201

Calpath: /Users/rmr/Dropbox/instruments/RAD/RAD_CALIBRATION/201_uri/2112

PLOT FILE NAME: radcalplot_201_2112

Reference voltage = 4070.0 millivolts (TP16)

PSPCAL

-1.0,	-123.33
-0.5,	-60.46
-0.1,	-10.97
0.0,	0.93
0.1,	13.24
0.2,	25.43
0.5,	63.12
1.0,	125.63
2.0,	250.37
4.0,	499.48
8.0,	998.15

PIRCAL

-2.0,	-1681.20
-1.0,	-837.70
-0.8,	-667.70
-0.4,	-333.20
-0.2,	-162.70
0.0,	4.40
0.2,	173.90
0.4,	339.60
0.6,	510.20
1.0,	846.40

Reference Rcase=32960.0

Reference Rdome=32850.0

caseR

5600,	589.000
10000,	944.000
15600,	1302.000
20000,	1531.000
25600,	1772.000
30000,	1932.000

domeR

5600,	590.000
10000,	946.000
15600,	1305.000
20000,	1534.000
25600,	1775.000
30000,	1934.000

==== CASE TEMPERATURE =====

Case Rref = 32960, Rref based on measurements of v_t = 33156. %Error = -0.6

Case fit : -9.647e-07 -2.312e-03 -3.535e-01

CASE THERMISTOR MILLIVOLTS

Meas	ADC	ADC-Corrected
------	-----	---------------

589.0	587.1	589.1
944.0	940.3	943.7
1302.0	1297.3	1302.3
1531.0	1524.8	1530.9
1772.0	1764.7	1772.1
1932.0	1923.5	1931.9

CASE THERMISTOR OHMS

CalR	Meas	ADC	ADC-corrected
5600	5577	5556	5578
10000	9953	9902	9949
15600	15504	15421	15508
20000	19875	19746	19873
25600	25416	25231	25419
30000	29784	29536	29780

CASE THERMISTOR DEG C

CalR	Meas	ADC	ADC-correc
39.96	40.07	40.17	40.07
25.00	25.12	25.24	25.13
14.28	14.43	14.55	14.42
8.58	8.72	8.87	8.72
3.08	3.24	3.40	3.23
-0.37	-0.21	-0.03	-0.21

==== DOME TEMPERATURE =====

Dome Rref = 32850, Rref based on measurements of v_t = 33067. %Error = -0.7

Dome fit : 1.271e-07, -3.467e-03, 1.116e+01

DOME THERMISTOR MILLIVOLTS

Meas	ADC	ADC-Corrected
590.0	599.1	590.0
946.0	954.0	946.0
1305.0	1312.1	1305.3
1534.0	1539.8	1533.6
1775.0	1780.4	1775.0
1934.0	1939.0	1934.1

DOME THERMISTOR OHMS

CalR	Meas	ADC	ADC-corrected
5600	5569	5670	5569
10000	9948	10057	9948
15600	15504	15629	15509
20000	19871	19991	19863
25600	25407	25545	25408
30000	29743	29890	29746

DOME THERMISTOR DEG C

CalR	Meas	ADC	ADC-correc
39.96	40.11	39.63	40.11
25.00	25.13	24.86	25.13
14.28	14.43	14.24	14.42

8.58	8.72	8.59	8.73
3.08	3.24	3.12	3.24
-0.37	-0.18	-0.29	-0.19

==== PSP THERMOPILE CIRCUIT =====

PSP Calibration Gain (g) = 124.61, Offset (o) = 1.1 millivolts

==== PIR THERMOPILE CIRCUIT =====

PIR Calibration Gain (g) = 842.32, Offset (o) = 4.5 millivolts

RAD201

=====

RAD201 SETUP COMMANDS





L : 10
k : 8.09e-6
K : 3.79e-6
A : 01
V : 4070
C 0 : 33156
C 1 : -9.647e-07
C 2 : -2.312e-03
C 3 : -3.535e-01
D 0 : 33067
D 1 : 1.271e-07
D 2 : -3.467e-03
D 3 : 1.116e+01
g : 124.61
o : 1.10
G : 842.32
O : 4.51

VERIFY CASE-DOME TEMPS

=====

OHMS YSI CASE DOME

5600 39.96 39.85 39.75
10000 25.00 24.94 24.97
15600 14.28 14.30 14.38
20000 8.58 8.63 8.73
25600 3.08 3.19 3.29
30000 -0.37 -0.22 -0.12





THE EPPLEY LABORATORY, INC.

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840
 Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

Calibration Certificate

Instrument: Precision Spectral Pyranometer, Model PSP, Serial Number 35149F3

Procedure: This pyranometer was compared in Eppley's Integrating Hemisphere according to procedures described in *ISO 9847 Section 5.3.1* and Technical Procedure, TP01 of The Eppley Laboratory, Inc.'s Quality Assurance Manual on Calibrations.

Transfer Standard: Eppley Standard Precision Pyranometer, Model SPP, Serial Number 37501F3

Results:

Sensitivity: $S = 8.16 \mu\text{V} / \text{Wm}^{-2}$

Uncertainty: $U_{95} = \pm 0.91\%$ (95% confidence level, $k=2$)


Resistance: 727Ω at 23°C

Date of Test: October 21, 2021

Traceability: This calibration is traceable to the World Radiation Reference (WRR) through comparisons with Eppley's AHF standard self-calibrating cavity pyrheliometers which participated in the Twelfth International Pyrheliometric Comparisons (IPC XII) at Davos, Switzerland in September-October 2015. Unless otherwise stated in the remarks section below or on the Sales Order, the results of this calibration are "AS FOUND / AS LEFT".

Due Date: Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy.

Customer: URI // RMR
 Narragansett, RI // Seattle, WA

In Charge of Test: 

Eppley SO: 66054

Date of Certificate: November 3, 2021

Remarks:

End of Report



THE EPPLEY LABORATORY, INC.

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840
 Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

Calibration Certificate

Instrument: Precision Infrared Radiometer, Model PIR, Serial Number 35116F3

Procedure: This pyrgeometer was compared against Eppley's Blackbody Calibration System under radiation intensities of approximately 350 Wm^{-2} with an average ambient temperature of 19°C according to procedures described in Technical Procedure, TP05 of The Eppley Laboratory, Inc.'s Quality Assurance Manual on Calibrations.

Transfer Standard: Eppley Precision Infrared Radiometer, Model PIR, Serial Number 32227F3


Results:
Sensitivity: $S = 3.80 \mu\text{V} / \text{Wm}^{-2}$
Uncertainty: $U_{95} = \pm 1.7\%$ (95% confidence level, $k=2$)
Resistance: 649Ω at 23°C

Date of Test: October 21, 2021

Traceability: This calibration is traceable to the International Practical Temperature Scale (IPTS). Additionally, transfer standard PIR #32227F3 provides traceability to the World Infrared Standard Group (WISG) of pyrgeometers housed at the Infrared Radiometry Section of the World Radiation Centre (WRC-IRS). Unless otherwise stated in the remarks section below or on the Sales Order, the results of this calibration are "AS FOUND / AS LEFT".

Due Date: Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy.

Customer: URI // RMR
 Narragansett, RI // Seattle, WA

In Charge of Test: 

Eppley SO: 66054

Date of Certificate: November 3, 2021

Remarks:

End of Report