



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

Pressure Test Certificate

Test Date: 2019-01-08

Description: SBE-5T Submersible Pump

Sensor Information:

Replaced the main piston "O"-Rings.

Model Number: SBE-5T

Serial Number: 3572

Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For: 15 Minutes

High Pressure Test: 10000 PSI Held For: 60 Minutes

Passed Test: True

Tested By: PS

High pressure is generally equal to the maximum depth rating of the instrument

