



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2025-09-09**

Description: **SBE-32 Carousel**

### Sensor Information:

Model Number: **SBE-32**

Serial Number: **0175**

### Pressure Test Protocol:

Low Pressure Test: **40**      PSI      Held For: **15**      Minutes

High Pressure Test: **40**      PSI      Held For: **15**      Minutes

Passed Test: **True**

Tested By: **DD**

